

تلفون:  
فاكس:

السادة ..... / ..... المحترمين

رقم الاستدراج (Q291-2017,18)

لوازم مختبرات

الموضوع: عروض اسعار

إشارة إلى الموضوع اعلاه الرجاء تزويدنا بأسعار اللوازم التالية بالشيكل شامل لضريبة القيمة المضافة والتوصيل للجامعة العربية الأمريكية:

الرقم	اسم الصنف	الوحدة	الكمية	السعر	اجمالي تكلفة البند
1	Clinical Centrifuge	عدد	5		
	المجموع الكلي				

شروط الاستدراج

- 1- يرجى العلم ان اخر موعد لتسليم العروض هو يوم (الاربعاء) الموافق 27/6/2018
- 2- الاسعار شاملة لضريبة القيمة المضافة.
- 3- تحديد مدة التوريد من تاريخ استلام امر الشراء، وفي حال عدم الالتزام يتم فرض غرامة تأخير.
- 4- ختم وتوقيع المورد المعتمد مع كتابة رقم الاستدراج واسم الشركة كما هو مسجل لدى الضريبة على الطرف.
- 5- لا تقبل العروض الا بالظرف المغلق.
- 6- للمراجعة والاستفسار بخصوص المواصفات يرجى الاتصال على هاتف 042418888 رقم داخلي 1486 او فاكس 042510972 أو بريد الكتروني pnp@aauj.edu.
- 7- للاطلاع على العطاءات والاستدراجات يرجى الدخول الى الرابط <http://www.aauj.edu/tenders>



مدير دائرة اللوازم و المشتريات  
أ. حسن رباحة

## Bench Top Centrifuge

- **Origin: USA, Europe.**
- **Bench top centrifuge.**
- **Brushless motor.**
- **Microprocessor control.**
- **Control and display:**
  - Digital display.
  - actual values are indicated During centrifugation.
  - Speed display: RPM x 100, input in steps of 100
  - RCF key: indication during run.
  - Time (t/min): 1- 99 min, continuous operation.
  - Impulse key: for short centrifugation operations.
- **Rotor:**
  - fixed angle
  - 8-12 tubes.
  - for routine medical laboratory applications
  - MAX Speed: 5500 RPM
- **Safety:**
  - Emergency lid lock release.
  - Imbalance switch-off.
- **Power: 220 V.**



❖ Kindly your offer will not be taken into our consideration if you don't fill below table.

Centrifuge	Yes/No	Comment
<b>Manufacture</b>	Manufacture	
<b>Model</b>	Model	
<b>Origin</b>	Origin	
<b>Delivery time</b>	weeks	
<b>3 years warranty</b>	Yes/No	
<b>Bench top centrifuge with rotor .</b>	Yes/No	
<b>Brushless motor.</b>	Yes/No	
<b>Microprocessor control.</b>	Yes/No	
<b>Control and display:</b>	Yes/No	
- Digital display.	Yes/No	
- actual values are indicated During centrifugation.	Yes/No	
- Speed display: RPM x 100, input in steps of 100	Yes/No	
- RCF key: indication during run.	Yes/No	
- Time (t/min): 1- 99 min, continuous operation.	Yes/No	
- Impulse key: for short centrifugation operations.	Yes/No	
<b>Rotor:</b>	Yes/No	
- fixed angle	Yes/No	
- 8-12 tubes .	Yes/No	
- 2 size of buckets : for routine medical laboratory apps	Yes/No	
- MAX Speed : 5500RPM	Yes/No	
<b>Safety :</b>	Yes/No	
- Emergency lid lock release .	Yes/No	
<b>Date of manufacture the machine not before 2017</b>		
<b>User's and service manual.</b>	Yes/No	
<b>Price includes installation and team training in site.</b>	Yes/No	
<b>10 years availability for spare parts</b>	Yes/No	